



July 11, 2018

Dear RNA Medical Customer,

A review of the CVC223 CO-oximeter Calibration Verification Controls Lot# 722333 has prompted us to adjust the assigned values for the AVOXimeter® 1000E, AVOXimeter® 4000 and GEM OPL analyzers. Please see the chart below for the corrected values.

**CVC223 Lot: 722333 Exp: 2019-03**

CVC223 Expected Values - AVOX1000e			
Level 1 Lot 75130	Mean	Min	Max
THb	4.5	3.9	5.1
O2Hb	30.4	25.8	35
Level 2 Lot 74856	Mean	Min	Max
THb	8	7.3	8.7
O2Hb	92.7	88.2	97.2
Level 3 Lot 74952	Mean	Min	Max
THb	13.1	12	14.2
O2Hb	83.8	79.5	88.1
Level 4 Lot 75047	Mean	Min	Max
THb	16.3	15	17.6
O2Hb	54.3	50	58.6
Level 5 75230	Mean	Min	Max
THb	21.9	19.7	24.1
O2Hb	3.3	-4.7	11.3

CVC223 Expected Values - AVOX4000 / GEM OPL			
Level 1 Lot 75130	Mean	Min	Max
THb	4.5	3.9	5.1
O2Hb	28.9	24.3	33.5
COHb	72.2	65.2	79.2
MetHb	0.4	-2.5	3.3
Level 2 Lot 74856	Mean	Min	Max
THb	7.9	7.2	8.6
O2Hb	93.7	89.2	98.2
COHb	6.1	2.1	10.1
MetHb	0.5	-2.4	3.4
Level 3 Lot 74952	Mean	Min	Max
THb	13.1	12	14.2
O2Hb	84.5	80.2	88.8
COHb	15.9	11.4	20.4
MetHb	0.2	-2.7	3.1
Level 4 Lot 75047	Mean	Min	Max
THb	16.2	14.9	17.5
O2Hb	53.6	49.3	57.9
COHb	47	41.7	52.3
MetHb	0.1	-2.8	3
Level 5 75230	Mean	Min	Max
THb	22.1	19.9	24.3
O2Hb	1.7	-6.3	9.7
COHb	ORL	-	-
MetHb	0.2	-2.7	3.1

Please do not hesitate to contact RNA Medical with any questions:

Phone: 978-772-9070 / 800-533-6162

Fax: 978-598-7918

E-Mail: [customerservice@rnamedical.com](mailto:customerservice@rnamedical.com)

BIOMKT-199  
Rev. A 07/2018