



May 10, 2017

Dear RNA Medical Customer,

Expected Values have been added to RNA Medical CVC223 CO-Oximeter Calibration Verification Controls for a new analyzer, the Instrumentation Laboratory GEM Premier 5000. Please find the Expected Value ranges for CVC223 lot 622332 below.

**CVC223 Lot: 622332 Exp: 2018-03**

	tHb g/dL		O2Hb %		COHb %		MetHb %	
	Mean	Range	Mean	Range	Mean	Range	Mean	Range
<b>Level 1</b>	4.8	4.1 - 5.5	29.5	26.5 - 32.5	70.7	68.7 - 72.7	-0.8	-2.8 - 1.2
<b>Level 2</b>	7.7	7.0 - 8.4	94.9	91.1 - 97.9	4.6	2.6 - 6.6	0.3	-1.7 - 2.3
<b>Level 3</b>	13.0	12.3 - 13.7	83.4	80.4 - 86.4	15.5	13.5 - 17.5	0.5	-1.5 - 2.5
<b>Level 4</b>	15.5	14.8 - 16.2	54.5	51.5 - 57.5	43.9	41.9 - 45.9	0.4	-1.6 - 2.4
<b>Level 5</b>	21.0	20.0 - 22.0	2.6	-0.4 - 5.6	94.7	92.7 - 96.7	1.3	-0.7- 3.3

Please do not hesitate to contact RNA Medical with any questions:

RNA Medical, Division of Bionostics, Inc.  
7 Jackson Road  
Devens, MA 01434  
Phone: 978-772-9070 / 800-533-6162  
Fax: 978-598-7918  
E-Mail: customerservice@rnamedical.com

BIOMKT-175  
Rev. A 05/2017